



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, BATHINDA  
JODHPUR ROMANA, MANDI DABWALI ROAD, BATHINDA, PUNJAB - 151001  
ਅਖਿਲ ਭਾਰਤੀ ਆਯੁਰਵਿਗਿਆਨ ਸੰਸਥਾਨ, ਬਠਿੰਡਾ ਅਖਿਲ ਭਾਰਤੀ ਆਯੁਰਵਿਗਿਆਨ ਸੰਸਥਾਨ, ਬਠਿੰਡਾ

**PROCUREMENT CELL**



No. AIIMS Bathinda/Procurement/NOC/1498

14 Dec 2022

Subject: Purchase of "Combined Energy Generator which activates both ultrasonic and bipolar energy simultaneously in a single unit" on Proprietary basis- **Inviting Comments thereon.**

The request received from Department of Obstetrics and Gynaecology, AIIMS Bathinda for the purchase of captioned Item from M/s Harish & Harish Surgical, 2-3, Camp Chowk, Delhi Road, Hisar, Haryana - 125001 authorized distributor of **M/s Olympus Corporation, Japan** - on Proprietary basis.

The Notice is being uploaded for general information of Aspirant Manufacturer/Dealer/Distributor to submit their objections/proposal, if any, on proprietorship of this item.

In case, the product of any Manufacturer/Authorized distributor/dealer conforms to the enclosed specifications, they may submit their proposal for the supply same Equipment along with the following: -

- Equipment brochure
- Point-by-Point compliance of the enclosed specifications, along with all relevant documentary evidence:

**The objection/proposal** should be sent in sealed cover to the **OIC, Procurement Cell, First Floor, C-Block, Admin. Building, AIIMS Bathinda, Mandi Dabwali Road, Bathinda - 151001 so as to reach on or before 26 Dec 2022 up to 12:00 PM.**, failing which it will be presumed that no any other vendor is interested to offer comments/protest and case will be decided accordingly on its **merit**.

The reference number should be super scribed on sealed envelope.

PROCUREMENT CELL  
AIIMS Bathinda

**Prof. (Dr.) Akhilesh Pathak**  
OIC, Procurement Cell

- Enclosed:**
1. PAC Certificate
  2. Authorization Letter
  3. Technical Specifications

1. **PAC Certificate**

**OLYMPUS**

Date: 01<sup>st</sup> June 2022

Ref No: OMSI/22-23/186

To,  
The Director  
All India Institute of Medical Sciences  
Bathinda, Punjab 151001

Sub: Proprietary Article Certificate "Combined Energy Generator which activates both ultrasonic and bipolar energy simultaneously in a single probe/unit" for department of Gynaecology.

Dear Sir,

We, M/s. Olympus Medical Systems India Private Limited having our registered office at Ground Floor, Tower-C, SAS Tower, The Medicity Complex, Sector-38, Gurgaon - 122 001, Haryana is subsidiary of Olympus Corporation of Asia Pacific Ltd. We further state that M/s. Olympus Corporation, Japan

This is to certify and confirm that the **"Combined Energy Generator which activates both ultrasonic and bipolar energy simultaneously in a single probe/unit and having under water cutting"** is being solely manufactured by M/s Olympus Corporation having their office at having their office at their office at 2951, Ishikawa-Cho, Hachioji-shi, Tokyo 192-8507, Japan and Olympus Winter & Ibe GmbH, P.O. Box - 701709, 22017 Keuhnstrabe 61122045 Hamburg Germany respectively.

This Product and feature are proprietary of Olympus Corporation, Japan. No company in the world is manufacturing such products as of now.

Thanking you,

For Olympus Medical Systems India Pvt. Ltd.

Authorized signatory

**OLYMPUS MEDICAL SYSTEMS INDIA PRIVATE LIMITED**

Ground Floor, Tower C, SAS Tower, The Medicity Complex, Sector - 38, Gurgaon, 122001, Haryana, INDIA  
Tel: 0124 4999193, Fax: 0124 4999190 | Web site: www.olympusmedical.com | CIN: U33110HR2009PT0009011

*[Handwritten signatures]*

2. Authorization Letter

**OLYMPUS**

AUTHORIZATION FORM

Date: 01<sup>st</sup> June 2022

Ref No: OMSI/22-23/185

To:  
The Director  
All India Institute of Medical Sciences  
Bathinda - Punjab-151001

Sub: Procurement for "Combined Energy Generator which activates both ultrasonic and bipolar energy simultaneously in a single unit and having under water cutting" for department of Gynaecology.

Whereas:

We M/s Olympus Medical Systems India Private Limited having our registered office at Ground Floor, Tower-C, SAS Tower, The Medicity Complex, Sector-38, Gurgaon - 122001, Haryana are subsidiary of M/s Olympus Corporation of Asia Pacific Ltd of M/s Olympus corporation Japan.

We further state that M/s Olympus Corporation, Japan and M/s Olympus Winter & IBE GMBH, Germany are established and reputable manufacturers of "Combined Energy Generator which activates both ultrasonic and bipolar energy simultaneously in a single unit and having under water cutting" having their office at their office at 2951, Ishikawa-Cho, Hachioji-shi, Tokyo 192-8507, Japan and Olympus Winter & Ibe GmbH, P.O. Box - 701709, 22017 Keuhnstrabe 61122045 Hamburg Germany respectively.

We do hereby authorize M/s Harish & Harish Surgical, having registered Office at 2-3, Camp Chowk Delhi Road, Hissar -125001 Haryana to submit tenders, and subsequently negotiate and sign the contract with you against this Proprietary tender of "Combined Energy Generator which activates both ultrasonic and bipolar energy simultaneously in a single unit and having under water cutting".

No company, or firm or individual other than M/s Harish & Harish Surgical, is authorized to bid, negotiate, and conclude the contract about this business against this specific subject matter. This authorization is valid for above subject matter only.


Yours faithfully,

For Olympus Medical Systems India Pvt. Ltd.

  
Authorized Signatory

**OLYMPUS MEDICAL SYSTEMS INDIA PRIVATE LIMITED**

Ground Floor, Tower-C, SAS Tower, The Medicity Complex, Sector-38, Gurgaon-122001, Haryana, INDIA  
Tel: 0124-4993111 Fax: 0124-4999190 | Website: www.olympusmedical.co.in | E-mail: (113110)HR2009@TC03@olj



### 3. Technical Specifications

#### Technical Specification

Technical Specification for "Combined energy generator" which activates both ultrasonic and bipolar energy simultaneously in a single probe/unit for department of Gynaecology.

1. The Device should provide two (02) types of energy Ultrasonic energy and (Bipolar energy delivered through single probe
2. The system should be capable of delivering Integrated Ultrasonic and RF Energy and dedicated advance RF energy from Single instrument for sealing & cutting vessels up to 7mm
3. System should have a universal connector for connecting Integrated Ultrasonic and Bipolar Energy instruments/ dedicated ultrasonic energy instruments.
4. The system should deliver advance RF Energy to seal artery, veins along with tissue bundles including and up to 7mm and fused vessels should be able to withstand more than 3 times of normal systolic blood pressure.
5. The system should be equipped with Unique Safety Assist Technology (Intelligence Tissue Monitoring) to prevent unintended Teflon damage of Probe.
6. System should have single autoclavable transducer for both Open & Laparoscopic Probes
7. There should not be any locking of the transducer and Transducer should guarantee 100 autoclave cycles
8. System should be able to power ultrasonic instruments at 47 KHZ frequency.
9. The Electrical energy generator or the combined generator should have provision to deliver separately the Monopolar Energy, Bipolar Energy, Advanced Bipolar energy capable of sealing blood vessels up to 7 mm in diameter and Under water cutting with both Monopolar and Bipolar Electrodes
10. The device should be functional for both Open and Laparoscopic Surgery
11. It should be compatible with probes that deliver ultrasonic energy and advanced bipolar energy either integrated in a single probe or separately but with a universal connect, if applicable
12. The probes should have jaw design which distributes pressure evenly for consistent sealing, should have both grasping and dissecting capabilities, should ensure rapid tissue cutting, minimal thermal spread, low smoke and mist generator for clearer surgical view.

DR. DEEPA KUMARI  
Professor & Head  
Dept. of Obstetrics & Gynecology  
AIIMS Bathinda

DR. SUPRITI KUMARI  
Assistant Professor  
Dept. of Obstetrics & Gynecology  
AIIMS Bathinda

*Deepa*  
*Supriti*

