

Tender Enquiry No.: AIIMS/BTI/Tender/47 for the Supply of Instruments for AV Fistula surgical Set and CAPD set

Pre-Bid Meeting was held on 31 May 2022 from 3 pm to 4 pm in Department of Nephrology, at AIIMS, Bathinda. The meeting was attended by Dr. Saurabh Nayak and Mr. Deepak Sharma (Ahuja Surgical). After listening to the queries raised by the vendors following changes have been made and to be uploaded as corrigendum.

Schedule No. / Item No.	Pere Bid Query	Corrigendum (Revised specs)
AVF Surgical instruments		
4	Mosquito Straight 12 cm versus 4 "	Size to be read as 12 cm
5	Mosquitoes Curved 12 cm versus 4 "	Size to be read as 12 cm
16	Whether Instrument container should be perforated, aluminium body or both	Perforated aluminium body should be there and dimensions 260 x 240 x 25 (mm)
17	Sponge holding Forcep 25 cm versus 18 cm	Size to be read as per 12 cm
21	Kidney tray 15 cm & 20 cm	Two kidney trays are required one of 15 cm and another of 20 cm
CAPD surgical set		
6	Kidney tray 15 cm & 25 cm	Two kidney trays are required one of 15cm and another of 25 cm
7	Mosquito Straight 12 cm versus 4 "	Size to be read as 12 cm
8	Mosquitoes Curved 12 cm versus 4 "	Size to be read as 12 cm

Dr. Saurabh Nayak


DR. SAURABH NAYAK

Assistant Professor
Dept. of Nephrology
AIIMS Bathinda

31/5/2022