



CORRIGENDUM

1. Refer tender reference No. AIIMS/BTI/Tender/73 for the procurement of Single Tender for Dual Energy X-Ray Absorptiometry (DEXA) on proprietary basis, the following changes in the technical specifications are hereby made:

(a) Stadiometer and Weighing Scale

For -

Imported digital stadiometer and electronic weighing scale with a minimum weight capacity of 200kg.

Read -

Stadiometer and electronic weighing scale with a minimum weight capacity of 200kg

(b) Computer Hardware

For -

iv. Color Laser Printer

Read -

iv. Mono Laser Printer

(c) Manpower

For-

Provision of the technician for the minimum first year of installation

Read-

Should provide Technical Assistance for minimum of first year of installation

(d) Other Requirements

For-

20 tonne Air Conditioning or Equivalent cooling to be provided by the vendor

Read-

Omitted

In addition to above amendments, the following technical points are also added to the technical specification of tender:

1. Quality Assurance should be done after every two years by the vendor in the warranty and CAMC Period.
2. As the item involves radiation exposure, thus, the unit should be AERB approved.
3. One revolving chair for technician and one stool for patient.

Dr. Rajesh Maurya
Assistant Professor
Burns and Plastic Surgery
AIIMS Bathinda.

[Signature]

DR. SAURABH NAYAK
Assistant Professor
Dept. of Nephrology
AIIMS Bathinda.

[Signature]

DR. MALTAF MIR
Associate Professor
Dept. of Burns and Plastic Surgery
AIIMS Bathinda

[Signature]
30/1/23.

DR. SHIVANI
Assistant Professor
Dept. of Endocrinology & Metabolism
AIIMS Bathinda