



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, BATHINDA
JODHPUR ROMANA, MANDI DABWALI ROAD, BATHINDA, PUNJAB- 151001
ਅਖਿਲ ਭਾਰਤੀ ਆਯੁਰ ਵਿਗਿਆਨ ਸੰਸਥਾਨ, ਬਠਿੰਡਾ ਅਖਿਲ ਭਾਰਤੀ ਆਯੁਰਵਿਗਿਆਨ ਸੰਸਥਾਨ, ਬਠਿੰਡਾ
Department of Obstetrics & Gynaecology



CORRIGENDUM

1. Refer tender reference No. AIIMS/BTI/Tender/77 with tender id 2023_AIIMB_742039_2 for the procurement of Cardiotocography (CTG) Machine for the Department of Obstetrics and Gynaecology at AIIMS Bathinda, the following changes in the technical specifications are hereby made:

(a) Product and Functional Requirements

For -

Toco and maternally sensed fetal movements are detected both manually

Read -

Wireless Toco and Maternally sensed fetal movements are detected both manually and automatically

(b) Product and Functional Requirements

For -

Ultrasound transducer for measuring fetal heart rate with auto correlation

Read -

Wireless Ultrasound transducer for measuring fetal heart rate with auto correlation

(c) Product and Functional Requirements

For-

Transducer should be waterproof

Read-

Wireless Transducer should be waterproof