ALL INDIA INSTITUTE OF MEDICAL SCIENCES, BATHINDA



Mandi Dabwali Road, Bathinda (Pb.) – 151001

PROCUREMENT CELL

AIIMS Bathinda/Proc.Cell/2025/ 172

Date: 14/07/2025 Due Date: 28/07/2025

Notice Inviting Quotations (NIQ)

The following items are required as per Annexure - A for AIIMS, Mandi Dabwali Road, Bathinda – 151001:

Sr.	Description	Specification	
i.	Consumables for CSSD	Annexure - A	

Kindly arrange to send your quotation giving the lowest rates per unit along with terms and conditions in a sealed cover addressed to the **Executive Director**, **AIIMS Bathinda**, **Mandi Dabwali Road**, **Bathinda – 151001**, **Punjab** so as to reach this office before the last date of receiving Quotations, i.e., on **28/07/2025**, **Time 17:00**

The word "Quotations of Consumables for CSSD for AllMS Bathinda, our reference number and the date should boldly be mentioned on the cover.

- 1. While submitting the quotation following should invariably be mentioned:
 - a) Details of Specification and its lowest rate F.O.R. AIIMS Bathinda.
 - b) GST Registration Number (attach copy)
 - c) PAN No. (Attach copy)
 - d) Clearly mentioned GST rates.
 - e) Tax at a concessional rate as applicable to the Educational Institutions
 - f) Completion from the date of receipt of the confirmed order.
- 2. The rates will be valid for one year from the date of issuance of the Work Order.
- 3. Payment will be made after the supply and inspection of the delivered items.
- 4. Under no circumstances will unsealed quotations be entertained in the office.
- 5. Quotations received after the due date shall not be considered.

1.9025 OIC, Procurement Cell

UNDERTAKING

(To be returned with quotation)

We hereby undertake the following:

1. We will not sell the product (s) to other institutions, bodies and also in the market at rates less than those quoted by us to AIIMS Bathinda.

2. The rate of GST/Excise Duty mentioned in the quotation is in accordance with the provisions of the rules, and the same is payable to the concerned authorities in respect of the stores.

3. The goods/Stores/articles offered shall be of the best quality and workmanship and their supply will be strictly in accordance with the technical specifications and particulars as detailed in the quotation.

4. The information furnished by us in the quotation is true and correct to the best of our knowledge and belief.

5. We have read and understood the rules, regulations, terms and conditions and agree to abide by them.

Authorised Signatory (Seal)

Quotation Proforma

(To be returned with Quotation)

Following proforma should be filled in and duly signed by the firm and sent along with the quotation.

Firm	's Reference				Date				
Firm Registration No. (if any)				PAN				-	
GST No.			Address:					З ₂	
Phone No.									
Email Id								T	
M/s:		i e de		NIQ No.					
				Last Date of Submission	f			a - 4	
Sr.	Product with Description Qt		Qty	Unit Price	GST Tota Unit Pric		t	Total Value	Total Value in words
1					×			di.	
2			·		1			AL CON	
3									
	Total		1940 -			1	9º sal	1.2784	in the second

I/we engage to supply the material(s) as required to your office and comply the following:

- i. Terms and conditions of the NIQ.
- ii. That the offer is valid for 90 days.
- iii. That we have not been debarred by any Government body/PSU/Autonomous Body
- iv. That the rates quoted are not higher than the rates quoted for same item to any Government body/PSU/Autonomous Body.
- v. That the quotation submitted by us is properly sealed and prepared so as to prevent any subsequent alteration and replacement.

Place: Date: Authorized Signatory (seal)

Terms and Conditions

 The quotation must be in the form furnished by Procuring Entity and should be free from corrections/erasures. In case there is any unavoidable correction it should be properly attested. If not, the quotation will not be considered. Quotation written in pencil will not be considered.
 All India Institute of Medical Quitable Correction Institute Constitute Constite Constitute Constitute Constitute Constitute Constitute Const

All India Institute of Medical Sciences (AIIMS), Bathinda reserves the right to accept the offer by individual items and reject any or all quotations without assigning any reason thereof and does not bind itself to accept the lowest quotations.
 The manufacturer's name and country for the lowest quotations.

The manufacturer's name and country of origin of materials offered must be clearly specified. Please quote whether your organisation is a large-scale industry or a small-scale industry. If you have NSIC/MSE/MSI/DGS&D Certificate, please attach it to the quotation. Mention your registration details.
 Complete details and USL energification of the statement of the stateme

4. Complete details and ISI specification if any must accompany the quotation. Make/brand of the item shall be stated wherever applicable. If you have got any counter offer as suitable to the material required by us, the same may be shown separately.

5. Samples must be submitted where specified along with the quotations. Samples must be carefully packed, sealed and labelled clearly with enquiry number, subject and sender's name for easy identification. Rejected samples will be returned at your cost if insisted.

 All supplies are subject to inspection and approval before acceptance. Manufacturer/supplier warranty certificates and manufacturer/Government approved lab test certificate shall be furnished along with the supply, wherever applicable.
 AIIMS Bathinda reserves the right to madify the supply.

AIIMS Bathinda reserves the right to modify the quantity specified in this enquiry.
 The prices quoted should be firm till the supplies are an initial.

8. The prices quoted should be firm till the supplies are completed. Please quote the rates in words and figures. Rates quoted should be free delivery at destination including all charges otherwise the quotation is likely to be rejected. Prices quoted for free delivery at destination will be given preference. If there is no indication regarding the FOR, in the quotation, then it will be considered as FOR destinations. Price quoted should be net and valid for a minimum period of 90 days from the date of opening of the quotation.

9. Delivery period required for supplying the material should be invariably specified in the quotation.

10. In case your quotation is accepted and order is placed on you, the supply against the order should be made within the period stipulated in the order. AIIMS Bathinda reserves the right to recover any loss sustained due to delayed delivery by way of penalty. Failure to supply the material within the stipulated period shall entitle Procuring Entity for the imposition of penalty without assigning any reasons @ 0.5% (half percent) of the total value of the item covered in order as penalty per day subject to a maximum of 5% (five percent) unless extension is obtained in writing from the office on valid ground before expiry of delivery period.

11. If the deliveries are not maintained and due to that account Procuring Entity is forced to buy the material at your risk and cost from elsewhere, the loss or damage that may be sustained there by will be recovered from the defaulting supplier.

12. Dispute clause: Any dispute relating to the enquiry shall be subject to the jurisdiction of the court at Bathinda only.

13. Our normal payment terms are 100% within 30 (thirty) days on receipt and acceptance of material at our site in good condition.

Sr.	Items Name	Qty.	Specification		
51.		uty.	Self-contained Biological Indicator for ETO		
	C.		sterilization processes.		
			 Its population should be 10^6 Spores/carrier 		
			 The Packaging is 100 Pcs in each box 		
			 The biological indicator label indicator color 		
1.	Biological ETO	100	change.		
			Should be compatible read-out time 48h		
			 Monitoring vacuum-assisted and gravity air- 		
	Diological Ere		displacement steam sterilization process at		
			121-135 ^O C		
			 Bacillus atrophaeus 10^6 spores per vial. Company Should be EN ISO 13485 certified 		
			Every box should contain Quality Assurance		
		at.	certificate with mention of Spore Population-		
			Value & Z Value.		
			 Self-contained Biological Indicator for Steam 		
	Biological Steam	100	sterilization processes.		
			Its population should be 10^6 Spores/carrier		
			 The Packaging is 100 Pcs in each box The biological indicator label indicator should be 		
			 The biological indicator label indicator should be color change turn to brown/black. 		
			 Should be compatible read-out time 24h 		
			 Monitoring vacuum-assisted and gravity air- 		
2.			displacement steam sterilization process at		
			121-135OC		
		*	Geobacillus stearothermophilus 10^6 spores		
			 per vial. Company Should be EN ISO 13485 certified 		
			Every box should contain Quality Assurance		
			certificate with mention of Spore Population-		
			Value & Z Value.		
			 label steam compatible with 3 Line labelling 		
			Gun		
	Lable Roll Steam	60	labels should be of minimum three-line label label should be beauty metal free and latex free		
			 label should be heavy metal free and latex free. It should have double adhesive with alpha 		
3.			numeric & numerical letters with 750 pieces per		
			roll.		
			Process indicator, after sterilization, turn to		
			black		
_			Should not leave any adhesive residue, for		
1	Street Mittain St.		Steam Sterilization cycles.		
aut	With the pile to	4	 Should be lead free Confirm to standard ISO 11140-1 class 1 		
	the fail the state of the				

Dr. Tarun Goyal Protessor & 11 ad Department of Orthopaedics AllMS Bathinda .

			 Company Should be EN ISO 13485 certified Should provide certificate of analysis/technical data sheet of each lot with Invoice. Manufacturer/importer should have valid manufacturing licence issued by CDSCO as under Class B Medical Device.
4.	Lable Roll ETO	60	 label ETO compatible with 3 Line labelling Gun labels should be of minimum three-line label label should be heavy metal free and latex free. It should have double adhesive with alpha numeric & numerical letters with 750 pieces per roll. Should not leave any adhesive residue, for ETO Sterilization cycles. Should be lead free Confirm to standard ISO 11140-1 class 1 Company Should be EN ISO 13485 certified Should provide certificate of analysis/technical data sheet of each lot with Invoice. Manufacturer/importer should have valid manufacturing licence issued by CDSCO as under Class B Medical Device
5.	Chemical Indicator Class 4 ETO	1000	 Class-4 eto indicators to monitor temperature, pressure & time. Color change indicator readout as per ISO 11140-1 It is Self-Adhesive. Type 4 Chemical Indicator, meeting ISO 11140-1 standards, providing reliable monitoring for Ethylene Oxide sterilization processes Company Should be EN ISO 13485 certified
6.	Chemical Indicator Class 5	1000	 Steam Chemical Integrator Strips contains no lead or other toxic heavy metals. Class-5 steam integrating indicators to monitor temperature, pressure & time. Color change indicator readout as per ISO 11140-1 It is Self-Adhesive. A color transition from pink to a shade of black provides indication of exposure to high temperature steam The Packaging is 200 Pcs in each box Company Should be EN ISO 13485 certified The supplier should have valid manufacturing license or import license from CDSCO as under Class B Medical Device
7.	5 cm Peel Pouch	15	Sealing line 3line sealing with printing on the
8.	10 cm Peel Pouch	15	sealing edge only.

Dr. Forvin Goyal Professor & Head Department of Orthopaedics AUMS Bathinda

6.

4

_			45	Minimum O II II III III III
	9.	15 cm Peel Pouch	15	 Minimum Sealing side should be >8mm The Standard law the bould be 200mtr
	10.	20 cm Peel Pouch	15	The Standard length should be 200mtr
-	11.	25 cm Peel Pouch	15	 It Should be suitable for both steam and ETO
	12.	30 cm Peel Pouch	15	sterilization methods.
	13.	35 cm Peel Pouch	15	 Heat-sealable sterilization reels should be constructed from transparent PET/PP multilayers copolymer film and medical grade paper. Process indicator for steam and ethylene oxide sterilization are applied on the paper surface of the reel and help to differentiate between Processed and unprocessed packages. Superior barrier with minimum 60gsm medical grade Paper conforming EN 868-3 & EN ISO 13485 Transparent, reinforced multilayer copolymer PET/PP film Three Independent seal lines Company Should be EN ISO 13485 certified Should provide certificate of analysis of each lot with Invoice. Shelf-life study report to be submit. Sterilization flat reels(steam) – 6 months Sterilization flat reels (ETO) – 1 YEAR
	14.	PCD Steam	100	 It should have strip. Type-2 chemical indicator to be used in process challenge device according to EN ISO-11140-1. It should be suitable for steam sterilization process. should have adhesive for documentation it should be lead free Color change should be turn to black for easy identification after processed. Helix device to be provided free of cost by the supplier Company Should be EN ISO 13485 certified Should provide certificate of analysis/technical data sheet of each lot with Invoice. Manufacturer/importer should have valid manufacturing licence issued by CDSCO as under Class B Medical Device
	15.	PCD ETO	100	 It should have strip. Type-2 chemical indicator to be used in process challenge device according to EN ISO-11140-1. It should be suitable for eo sterilization process. should have adhesive for documentation it should be lead free Color change should be turn to ref. colour or equivalent for easy identification after processed.

Dr. Tarun Goyal Professor & Head Peptriment of Orthopaedics Department of Orthopaedics AllMS Bathinda 4 U

			 Helix device to be provided free of cost by the supplier Company Should be EN ISO 13485 certified Should provide certificate of analysis/technical data sheet of each lot with Invoice. Manufacturer/importer should have valid manufacturing licence issued by CDSCO as under Class B Medical Device
16.	Bowie & Dick Test	480	 Bowie dick test packs for monitoring Vacuum formation in Autoclave chambers. Its along with penetration challenge material should be in a single easy handled pack. Should be lead free. Weight per pack should be 280 – 300 gms Sheet quantity 210 sheets on each side (+ - 5) to given enough challenge to Sterilizer Bowie dick test sheet color change should be turn to Black after processed Company Should be EN ISO 13485 certified Should provide COA/declaration of conformity of each lot with Invoice. Manufacturer/importer should have valid manufacturing licence issued by CDSCO as under Class B Medical Device.

